

Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\FID_F\Data\FF110118\
 Data File : FF002550.D
 Signal(s) : FID2B.ch
 Acq On : 01 Nov 2018 18:40
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 02 00:56:35 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF101218.M
 Quant Title :
 QLast Update : Fri Oct 12 13:21:40 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.653	6203433	44.632 ug/ml
Target Compounds			
2) N-DECANE	4.378	8042414	52.649 ug/ml
3) N-DODECANE	6.400	8057970	52.039 ug/ml
4) N-TETRADECANE	8.194	7994924	51.091 ug/ml
5) N-HEXADECANE	9.796	7893325	50.222 ug/ml
6) N-OCTADECANE	11.239	7744262	48.729 ug/ml
7) N-EICOSANE	12.551	7465858	47.137 ug/ml
8) N-DOCOSANE	13.751	7188701	45.566 ug/ml
10) N-TETRACOSANE	14.857	6917333	44.214 ug/ml
11) N-HEXACOSANE	15.881	6756006	43.387 ug/ml
12) N-OCTACOSANE	16.833	6722686	43.340 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\FID_F\Data\FF110118\
Data File : FF002550.D
Signal(s) : FID2B.ch
Acq On : 01 Nov 2018 18:40
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 02 00:56:35 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF101218.M
Quant Title :
QLast Update : Fri Oct 12 13:21:40 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

