

rters

Instrument :
FID_F
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Ankita Jodhani 11/02/2021
Supervised By :mohammad ahmed 11/02/2021

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF11012
Data File : FF009245.D
Signal(s) : FID2B.ch
Acq On : 01 Nov 2021 22:39
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100721.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.360	2.340	2.410	BB	785202	5752468	99.36%	6.135%
2	4.466	4.422	4.514	PB	656112	5789446	100.00%	6.175%
3	6.484	6.455	6.540	BB	589297	5720908	98.82%	6.101%
4	8.276	8.246	8.342	BB	544511	5620883	97.09%	5.995%
5	9.873	9.836	9.931	BB	542485	5595614	96.65%	5.968%
6	11.312	11.272	11.365	BB	507401	5604149	96.80%	5.977%
7	12.620	12.577	12.670	BB	480279	5495943	94.93%	5.862%
8	13.815	13.770	13.867	BB	438645	5313800	91.78%	5.667%
9	14.710	14.660	14.770	BB	330799	4595810	79.38%	4.902%
10	14.916	14.865	14.971	BB	428715	5237111	90.46%	5.585%
11	15.935	15.883	15.990	BB	400738	5082471	87.79%	5.421%
12	16.882	16.836	16.934	BB	379842	4950897	85.52%	5.280%
13	17.767	17.717	17.819	BB	361037	4866473	84.06%	5.190%
14	18.598	18.538	18.642	BB	359984	4779939	82.56%	5.098%
15	19.379	19.113	19.431	BB	338481	4665335	80.58%	4.976%
16	20.118	19.770	20.152	BV	326022	4708981	81.34%	5.022%
17	20.855	20.712	20.969	BB	270131	4949299	85.49%	5.279%
18	21.771	21.684	21.901	VBA	199513	5033152	86.94%	5.368%
Sum of corrected areas:						93762681		

FF100721.M Tue Nov 02 09:38:20 2021