

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF110724AL\
 Data File : FF014951.D
 Signal(s) : FID2B.ch
 Acq On : 07 Nov 2024 11:37
 Operator : YP\AJ
 Sample : P4708-02DL 5X
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 OR-02-110424-E2DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/07/2024
 Supervised By :Ankita Jodhani 11/08/2024

Integration File: sample.E
 Quant Time: Nov 07 12:00:36 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\Aliphatic EPH 102124.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 22 05:46:44 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.935	959610	6.293 ug/mlm
Spiked Amount 50.000		Recovery =	12.59%
12) S 1-chlorooctadecane (S...	13.373	949610	7.904 ug/mlm
Spiked Amount 50.000		Recovery =	15.81%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF110724AL\
Data File : FF014951.D
Signal(s) : FID2B.ch
Acq On : 07 Nov 2024 11:37
Operator : YP\AJ
Sample : P4708-02DL 5X
Misc :
ALS Vial : 19 Sample Multiplier: 1


Instrument :
FID_F
ClientSampleId :
OR-02-110424-E2DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/07/2024
Supervised By :Ankita Jodhani 11/08/2024

Integration File: sample.E
Quant Time: Nov 07 12:00:36 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\Aliphatic EPH 102124.M
Quant Title : GC Extractables
QLast Update : Tue Oct 22 05:46:44 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.935 min
Delta R.T.: 0.022 min
Response: 959610
Conc: 6.29 ug/ml

Instrument : FID_F
ClientSampleId : OR-02-110424-E2DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/07/2024
Supervised By :Ankita Jodhani 11/08/2024



#12 1-chlorooctadecane (SURR)

R.T.: 13.373 min
Delta R.T.: 0.024 min
Response: 949610
Conc: 7.90 ug/ml m