

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF110822\
 Data File : FF011556.D
 Signal(s) : FID2B.ch
 Acq On : 08 Nov 2022 16:43
 Operator : YP\AJ
 Sample : PB148882BS
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 PB148882BS

Integration File: autoint1.e
 Quant Time: Nov 09 04:59:05 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF110422.M
 Quant Title :
 QLast Update : Fri Nov 04 14:11:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.046	2484211	20.938 ug/ml
Target Compounds			
2) N-DECANE	4.611	1396209	10.579 ug/ml
3) N-DODECANE	6.770	1425183	10.604 ug/ml
4) N-TETRADECANE	8.593	1427988	10.552 ug/ml
5) N-HEXADECANE	10.198	1447093	10.526 ug/ml
6) N-OCTADECANE	11.639	1460237	10.557 ug/ml
7) N-EICOSANE	12.949	1428271	10.327 ug/ml
8) N-DOCOSANE	14.146	1408643	10.436 ug/ml
10) N-TETRACOSANE	15.249	1399837	10.322 ug/ml
11) N-HEXACOSANE	16.269	1377468	10.310 ug/ml
12) N-OCTACOSANE	17.218	1341764	10.128 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF110822\
Data File : FF011556.D
Signal(s) : FID2B.ch
Acq On : 08 Nov 2022 16:43
Operator : YP\AJ
Sample : PB148882BS
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
PB148882BS

Integration File: autoint1.e
Quant Time: Nov 09 04:59:05 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF110422.M
Quant Title :
QLast Update : Fri Nov 04 14:11:34 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

