

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF110822\
 Data File : FF011572.D
 Signal(s) : FID2B.ch
 Acq On : 09 Nov 2022 02:03
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Nov 09 05:01:06 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF110422.M
 Quant Title :
 QLast Update : Fri Nov 04 14:11:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.056	6273136	52.873 ug/ml
Target Compounds			
2) N-DECANE	4.617	7343186	55.636 ug/ml
3) N-DODECANE	6.778	7487272	55.707 ug/ml
4) N-TETRADECANE	8.602	7506677	55.470 ug/ml
5) N-HEXADECANE	10.208	7569564	55.061 ug/ml
6) N-OCTADECANE	11.650	7554038	54.611 ug/ml
7) N-EICOSANE	12.959	7464419	53.973 ug/ml
8) N-DOCOSANE	14.157	7206537	53.391 ug/ml
10) N-TETRACOSANE	15.259	7105877	52.399 ug/ml
11) N-HEXACOSANE	16.279	6912882	51.741 ug/ml
12) N-OCTACOSANE	17.229	6769542	51.099 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF110822\
Data File : FF011572.D
Signal(s) : FID2B.ch
Acq On : 09 Nov 2022 02:03
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Nov 09 05:01:06 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF110422.M
Quant Title :
QLast Update : Fri Nov 04 14:11:34 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

