

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF110921\
 Data File : FF009305.D
 Signal(s) : FID2B.ch
 Acq On : 09 Nov 2021 13:14
 Operator : AJ\MA
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 5 TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 11/10/2021
 Supervised By :mohammad ahmed 11/10/2021

Integration File: autoint1.e
 Quant Time: Nov 09 14:21:08 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF110921.M
 Quant Title :
 QLast Update : Tue Nov 09 14:19:53 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.705	527136	4.455 ug/ml
Target Compounds			
1) N-OCTANE	2.364	690495	4.725 ug/ml
2) N-DECANE	4.459	684690	4.616 ug/ml
3) N-DODECANE	6.476	639850	4.303 ug/ml
4) N-TETRADECANE	8.273	583547	3.960 ug/ml
5) N-HEXADECANE	9.869	639032	4.382 ug/ml
6) N-OCTADECANE	11.307	658090	4.522 ug/ml
7) N-EICOSANE	12.612	651248	4.564 ug/ml
8) N-DOCOSANE	13.807	633846	4.607 ug/ml
10) N-TETRACOSANE	14.908	620964	4.632 ug/ml
11) N-HEXACOSANE	15.927	590651	4.609 ug/ml
12) N-OCTACOSANE	16.874	565786	4.655 ug/ml
13) N-TRIACONTANE	17.760	542884	4.714 ug/ml
14) N-DOTRIACONTANE	18.588	562118	5.228 ug/ml
15) N-TETRATRIACONTANE	19.369	479869	4.794 ug/ml
16) N-HEXATRIACONTANE	20.108	635701	6.722 ug/mlm
17) N-OCTATRIACONTANE	20.844	395446	4.508 ug/ml
18) N-TETRACONTANE	21.765	546592	5.952 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF110921\
Data File : FF009305.D
Signal(s) : FID2B.ch
Acq On : 09 Nov 2021 13:14
Operator : AJ\MA
Sample : 5 TRPH STD
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
5 TRPH STD

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 11/10/2021
Supervised By :mohammad ahmed 11/10/2021

Integration File: autoint1.e
Quant Time: Nov 09 14:21:08 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF110921.M
Quant Title :
QLast Update : Tue Nov 09 14:19:53 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

