

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF111122\
 Data File : FF011684.D
 Signal(s) : FID2B.ch
 Acq On : 12 Nov 2022 17:08
 Operator : YP\AJ
 Sample : N5538-01MSD
 Misc :
 ALS Vial : 86 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 TP-14-A-WESTMSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/14/2022
 Supervised By :Ankita Jodhani 11/14/2022

Integration File: Sample.e
 Quant Time: Nov 14 02:31:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF110422.M
 Quant Title :
 QLast Update : Fri Nov 04 14:11:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.054	1605278	13.530 ug/ml
Target Compounds			
1) N-OCTANE	2.090	1829346	14.311 ug/ml
2) N-DECANE	4.612	2216499	16.794 ug/ml
3) N-DODECANE	6.773	2448602	18.218 ug/ml
4) N-TETRADECANE	8.598	2667722	19.713 ug/mlm
5) N-HEXADECANE	10.205	2702892	19.661 ug/mlm
6) N-OCTADECANE	11.648	2698447	19.508 ug/mlm
7) N-EICOSANE	12.957	2707872	19.580 ug/mlm
8) N-DOCOSANE	14.156	2428791	17.994 ug/ml
10) N-TETRACOSANE	15.259	3338520	24.618 ug/ml
11) N-HEXACOSANE	16.280	2440475	18.266 ug/ml
12) N-OCTACOSANE	17.231	2878461	21.727 ug/mlm
13) N-TRIACONTANE	18.118	2406157	18.244 ug/ml
14) N-DOTRIACONTANE	18.950	2137443	15.880 ug/ml
15) N-TETRATRIACONTANE	19.733	2289193	17.382 ug/ml
16) N-HEXATRIACONTANE	20.474	2421056	19.656 ug/mlm
17) N-OCTATRIACONTANE	21.217	2556641	21.958 ug/ml
18) N-TETRACONTANE	22.143	2943823	27.070 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF111122\
Data File : FF011684.D
Signal(s) : FID2B.ch
Acq On : 12 Nov 2022 17:08
Operator : YP\AJ
Sample : N5538-01MSD
Misc :
ALS Vial : 86 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
TP-14-A-WESTMSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/14/2022
Supervised By :Ankita Jodhani 11/14/2022

Integration File: Sample.e
Quant Time: Nov 14 02:31:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF110422.M
Quant Title :
QLast Update : Fri Nov 04 14:11:34 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

