

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF111418\
 Data File : FF002693.D
 Signal(s) : FID2B.ch
 Acq On : 14 Nov 2018 19:32
 Operator : AJ\SJ
 Sample : J5925-10
 Misc :
 ALS Vial : 62 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 15 00:50:42 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111418.M
 Quant Title :
 QLast Update : Wed Nov 14 13:46:00 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.646	981414	12.564 ug/ml
Target Compounds			
4) N-TETRADECANE	8.244f	362928	3.142 ug/ml
7) N-EICOSANE	12.594	137077	1.416 ug/ml
10) N-TETRACOSANE	14.814	80362	0.902 ug/ml
12) N-OCTACOSANE	16.822	103074	1.081 ug/ml
13) N-TRIACONTANE	17.713	149675	1.510 ug/ml
14) N-DOTRIACONTANE	18.549	153704	1.553 ug/ml
18) N-TETRACONTANE	21.759	482424	5.507 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF111418\
Data File : FF002693.D
Signal(s) : FID2B.ch
Acq On : 14 Nov 2018 19:32
Operator : AJ\SJ
Sample : J5925-10
Misc :
ALS Vial : 62 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 15 00:50:42 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111418.M
Quant Title :
QLast Update : Wed Nov 14 13:46:00 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#4 N-TETRADECANE

R.T.: 8.244 min
Delta R.T.: 0.055 min
Response: 362928
Conc: 3.14 ug/ml




#7 N-EICOSANE

R.T.: 12.594 min
Delta R.T.: 0.045 min
Response: 137077
Conc: 1.42 ug/ml




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.646 min
Delta R.T.: -0.003 min
Response: 981414
Conc: 12.56 ug/ml




#10 N-TETRACOSANE

R.T.: 14.814 min
Delta R.T.: -0.039 min
Response: 80362
Conc: 0.90 ug/ml




#12 N-OCTACOSANE

R.T.: 16.822 min
Delta R.T.: -0.005 min
Response: 103074
Conc: 1.08 ug/ml




#13 N-TRIACONTANE

R.T.: 17.713 min
Delta R.T.: -0.004 min
Response: 149675
Conc: 1.51 ug/ml



#14 N-DOTRIACONTANE

R.T.: 18.549 min
Delta R.T.: -0.002 min
Response: 153704
Conc: 1.55 ug/ml



#18 N-TETRACONTANE

R.T.: 21.759 min
Delta R.T.: 0.036 min
Response: 482424
Conc: 5.51 ug/ml