

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF111418\
 Data File : FF002680.D
 Signal(s) : FID2B.ch
 Acq On : 14 Nov 2018 12:52
 Operator : AJ\SJ
 Sample : 10 TRPH STD
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: Nov 14 13:20:07 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111418.M
 Quant Title :
 QLast Update : Wed Nov 14 12:37:49 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.645	1029512	12.496 ug/ml
Target Compounds			
1) N-OCTANE	2.267	1701710	12.983 ug/ml
2) N-DECANE	4.362	1667191	12.633 ug/ml
3) N-DODECANE	6.388	1619272	12.613 ug/ml
4) N-TETRADECANE	8.184	1549237	12.779 ug/ml
5) N-HEXADECANE	9.787	1454799	12.829 ug/ml
6) N-OCTADECANE	11.231	1386802	12.898 ug/ml
7) N-EICOSANE	12.544	1281692	12.737 ug/ml
8) N-DOCOSANE	13.744	1213656	12.642 ug/ml
10) N-TETRACOSANE	14.849	1171254	12.475 ug/ml
11) N-HEXACOSANE	15.873	1169599	12.182 ug/ml
12) N-OCTACOSANE	16.825	1212433	12.005 ug/ml
13) N-TRIACONTANE	17.714	1295429	12.291 ug/ml
14) N-DOTRIACONTANE	18.549	1347207	12.678 ug/ml
15) N-TETRATRIACONTANE	19.336	1297685	12.336 ug/ml
16) N-HEXATRIACONTANE	20.078	1252278	12.271 ug/mlm
17) N-OCTATRIACONTANE	20.819	1132264	12.080 ug/ml
18) N-TETRACONTANE	21.721	1229186	12.555 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF111418\
Data File : FF002680.D
Signal(s) : FID2B.ch
Acq On : 14 Nov 2018 12:52
Operator : AJ\SJ
Sample : 10 TRPH STD
Misc :
ALS Vial : 56 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: Nov 14 13:20:07 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111418.M
Quant Title :
QLast Update : Wed Nov 14 12:37:49 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

