

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF111418\
 Data File : FF002684.D
 Signal(s) : FID2B.ch
 Acq On : 14 Nov 2018 15:00
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: Nov 15 00:48:57 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111418.M
 Quant Title :
 QLast Update : Wed Nov 14 13:46:00 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.648	4117354	52.708 ug/ml
Target Compounds			
2) N-DECANE	4.370	6612643	52.944 ug/ml
3) N-DODECANE	6.395	6539295	53.952 ug/ml
4) N-TETRADECANE	8.191	6312121	54.643 ug/ml
5) N-HEXADECANE	9.793	5998650	54.863 ug/ml
6) N-OCTADECANE	11.236	5684311	54.759 ug/ml
7) N-EICOSANE	12.547	5264649	54.374 ug/ml
8) N-DOCOSANE	13.749	4897722	53.494 ug/ml
10) N-TETRACOSANE	14.852	4654078	52.241 ug/ml
11) N-HEXACOSANE	15.876	4611940	50.819 ug/ml
12) N-OCTACOSANE	16.827	4815111	50.484 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF111418\
Data File : FF002684.D
Signal(s) : FID2B.ch
Acq On : 14 Nov 2018 15:00
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: Nov 15 00:48:57 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111418.M
Quant Title :
QLast Update : Wed Nov 14 13:46:00 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

