

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF111723\
 Data File : FF013736.D
 Signal(s) : FID2B.ch
 Acq On : 17 Nov 2023 17:43
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 17 21:32:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111023.M
 Quant Title :
 QLast Update : Fri Nov 10 12:42:28 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.967	5700205	49.886 ug/ml
Target Compounds				
2)	N-DECANE	4.518	5895806	48.692 ug/ml
3)	N-DODECANE	6.683	6087243	49.286 ug/ml
4)	N-TETRADECANE	8.509	6121301	49.549 ug/ml
5)	N-HEXADECANE	10.115	6260841	50.192 ug/ml
6)	N-OCTADECANE	11.557	6306258	50.612 ug/ml
7)	N-EICOSANE	12.868	6303240	50.589 ug/ml
8)	N-DOCOSANE	14.067	6095517	50.177 ug/ml
10)	N-TETRACOSANE	15.171	6062876	50.196 ug/ml
11)	N-HEXACOSANE	16.192	5910498	50.085 ug/ml
12)	N-OCTACOSANE	17.142	5848807	50.824 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF111723\
Data File : FF013736.D
Signal(s) : FID2B.ch
Acq On : 17 Nov 2023 17:43
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 17 21:32:45 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111023.M
Quant Title :
QLast Update : Fri Nov 10 12:42:28 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

