

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF112018\
 Data File : FF002756.D
 Signal(s) : FID2B.ch
 Acq On : 20 Nov 2018 15:08
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 20 15:33:39 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111418.M
 Quant Title :
 QLast Update : Wed Nov 14 13:46:00 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.652	4490515	57.485 ug/ml
Target Compounds			
2) N-DECANE	4.375	5699473	45.633 ug/ml
3) N-DODECANE	6.396	5568959	45.947 ug/ml
4) N-TETRADECANE	8.191	5410355	46.836 ug/ml
5) N-HEXADECANE	9.792	5274882	48.243 ug/ml
6) N-OCTADECANE	11.237	5196899	50.064 ug/ml
7) N-EICOSANE	12.549	5111202	52.789 ug/ml
8) N-DOCOSANE	13.750	5073386	55.412 ug/ml
10) N-TETRACOSANE	14.855	5066574	56.871 ug/ml
11) N-HEXACOSANE	15.879	5119230	56.408 ug/ml
12) N-OCTACOSANE	16.831	5210377	54.628 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF112018\
Data File : FF002756.D
Signal(s) : FID2B.ch
Acq On : 20 Nov 2018 15:08
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 20 15:33:39 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111418.M
Quant Title :
QLast Update : Wed Nov 14 13:46:00 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

