

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF112125\
 Data File : FF016779.D
 Signal(s) : FID2B.ch
 Acq On : 21 Nov 2025 12:34
 Operator : YP\AJ
 Sample : PB170656BS
 Misc :
 ALS Vial : 72 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 PB170656BS

Integration File: autoint1.e
 Quant Time: Nov 25 08:51:36 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF102925.M
 Quant Title :
 QLast Update : Wed Oct 29 12:48:52 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.045	1983786	18.988 ug/ml
Target Compounds			
2) N-DECANE	4.564	2097503	18.957 ug/ml
3) N-DODECANE	6.743	2185251	19.672 ug/ml
4) N-TETRADECANE	8.575	2209707	19.780 ug/ml
5) N-HEXADECANE	10.184	2251895	20.042 ug/ml
6) N-OCTADECANE	11.629	2292490	19.589 ug/ml
7) N-EICOSANE	12.942	2269212	18.584 ug/ml
8) N-DOCOSANE	14.143	2254558	18.913 ug/ml
10) N-TETRACOSANE	15.250	2186699	18.455 ug/ml
11) N-HEXACOSANE	16.274	2126586	18.511 ug/ml
12) N-OCTACOSANE	17.227	2082022	18.270 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF112125\
Data File : FF016779.D
Signal(s) : FID2B.ch
Acq On : 21 Nov 2025 12:34
Operator : YP\AJ
Sample : PB170656BS
Misc :
ALS Vial : 72 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
PB170656BS

Integration File: autoint1.e
Quant Time: Nov 25 08:51:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF102925.M
Quant Title :
QLast Update : Wed Oct 29 12:48:52 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

