

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF112223\
 Data File : FF013754.D
 Signal(s) : FID2B.ch
 Acq On : 22 Nov 2023 11:50
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 22 13:28:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111023.M
 Quant Title :
 QLast Update : Fri Nov 10 12:42:28 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.966	6100326	53.387 ug/ml
Target Compounds			
2) N-DECANE	4.510	6333049	52.303 ug/ml
3) N-DODECANE	6.678	6509167	52.703 ug/ml
4) N-TETRADECANE	8.505	6556072	53.069 ug/ml
5) N-HEXADECANE	10.112	6717005	53.849 ug/ml
6) N-OCTADECANE	11.556	6763431	54.281 ug/ml
7) N-EICOSANE	12.867	6743486	54.122 ug/ml
8) N-DOCOSANE	14.066	6509099	53.582 ug/ml
10) N-TETRACOSANE	15.171	6452564	53.422 ug/ml
11) N-HEXACOSANE	16.192	6263684	53.077 ug/ml
12) N-OCTACOSANE	17.142	6180807	53.709 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF112223\
Data File : FF013754.D
Signal(s) : FID2B.ch
Acq On : 22 Nov 2023 11:50
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 22 13:28:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111023.M
Quant Title :
QLast Update : Fri Nov 10 12:42:28 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

