

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF112223\
 Data File : FF013760.D
 Signal(s) : FID2B.ch
 Acq On : 22 Nov 2023 14:45
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 22 22:48:53 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111023.M
 Quant Title :
 QLast Update : Fri Nov 10 12:42:28 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.967	6116695	53.531 ug/ml
Target Compounds			
1) N-OCTANE	1.997	5664836	47.054 ug/ml
2) N-DECANE	4.517	6316133	52.163 ug/ml
3) N-DODECANE	6.682	6496652	52.601 ug/ml
4) N-TETRADECANE	8.508	6524329	52.812 ug/ml
5) N-HEXADECANE	10.115	6711555	53.805 ug/ml
6) N-OCTADECANE	11.558	6777255	54.392 ug/ml
7) N-EICOSANE	12.868	6777087	54.392 ug/ml
8) N-DOCOSANE	14.067	6544032	53.869 ug/ml
10) N-TETRACOSANE	15.172	6509956	53.897 ug/ml
11) N-HEXACOSANE	16.192	6341369	53.736 ug/ml
12) N-OCTACOSANE	17.143	6287578	54.637 ug/ml
13) N-TRIACONTANE	18.031	6332614	56.196 ug/ml
14) N-DOTRIACONTANE	18.864	6150578	56.787 ug/ml
15) N-TETRATRIACONTANE	19.647	5256126	56.361 ug/ml
16) N-HEXATRIACONTANE	20.386	4141177	56.641 ug/ml
17) N-OCTATRIACONTANE	21.114	3724730	56.608 ug/ml
18) N-TETRACONTANE	22.002	3570748	56.756 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF112223\
Data File : FF013760.D
Signal(s) : FID2B.ch
Acq On : 22 Nov 2023 14:45
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 22 22:48:53 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111023.M
Quant Title :
QLast Update : Fri Nov 10 12:42:28 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

