

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF112223\
 Data File : FF013773.D
 Signal(s) : FID2B.ch
 Acq On : 22 Nov 2023 21:06
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 22 22:50:55 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111023.M
 Quant Title :
 QLast Update : Fri Nov 10 12:42:28 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.979	6106446	53.441 ug/ml
Target Compounds			
1) N-OCTANE	1.999	5733055	47.620 ug/ml
2) N-DECANE	4.523	6510055	53.765 ug/ml
3) N-DODECANE	6.689	6739011	54.564 ug/ml
4) N-TETRADECANE	8.517	6827550	55.266 ug/ml
5) N-HEXADECANE	10.124	6985151	55.999 ug/ml
6) N-OCTADECANE	11.568	6982389	56.038 ug/ml
7) N-EICOSANE	12.879	6910126	55.459 ug/ml
8) N-DOCOSANE	14.078	6612398	54.432 ug/ml
10) N-TETRACOSANE	15.183	6550400	54.232 ug/ml
11) N-HEXACOSANE	16.205	6402176	54.251 ug/ml
12) N-OCTACOSANE	17.156	6419352	55.782 ug/ml
13) N-TRIACONTANE	18.044	6578903	58.382 ug/ml
14) N-DOTRIACONTANE	18.877	6538221	60.366 ug/ml
15) N-TETRATRIACONTANE	19.661	5697143	61.090 ug/ml
16) N-HEXATRIACONTANE	20.400	4628114	63.301 ug/mlm
17) N-OCTATRIACONTANE	21.134	4176793	63.478 ug/ml
18) N-TETRACONTANE	22.029	4617094	73.387 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF112223\
Data File : FF013773.D
Signal(s) : FID2B.ch
Acq On : 22 Nov 2023 21:06
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 22 22:50:55 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111023.M
Quant Title :
QLast Update : Fri Nov 10 12:42:28 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

