

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_F\Data\FF112520\  
 Data File : FF007919.D  
 Signal(s) : FID2B.ch  
 Acq On : 25 Nov 2020 18:07  
 Operator : DD\AJ  
 Sample : L4880-06MSD  
 Misc :  
 ALS Vial : 63 Sample Multiplier: 1

Instrument :  
 FID\_F  
 ClientSampleId :  
 KENS-B4-112320-0-2MSD

Integration File: autoint1.e  
 Quant Time: Nov 26 00:33:32 2020  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_F\Method\FF110220.M  
 Quant Title :  
 QLast Update : Tue Nov 03 01:29:25 2020  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1uL  
 Signal Phase : Rxi-1ms  
 Signal Info : 20mx0.18mmx0.18um

| Compound                      | R.T.   | Response | Conc Units   |
|-------------------------------|--------|----------|--------------|
| -----                         |        |          |              |
| System Monitoring Compounds   |        |          |              |
| 9) S TETRACOSANE-d50 (SURR... | 14.711 | 1930542  | 15.114 ug/ml |
| Target Compounds              |        |          |              |
| 2) N-DECANE                   | 4.433  | 3032485  | 22.288 ug/ml |
| 3) N-DODECANE                 | 6.456  | 3083100  | 22.274 ug/ml |
| 4) N-TETRADECANE              | 8.253  | 3107677  | 22.007 ug/ml |
| 5) N-HEXADECANE               | 9.854  | 3054143  | 21.014 ug/ml |
| 6) N-OCTADECANE               | 11.298 | 3049942  | 20.551 ug/ml |
| 7) N-EICOSANE                 | 12.611 | 3018325  | 20.363 ug/ml |
| 8) N-DOCOSANE                 | 13.812 | 2928129  | 20.098 ug/ml |
| 10) N-TETRACOSANE             | 14.916 | 2818113  | 19.487 ug/ml |
| 11) N-HEXACOSANE              | 15.939 | 2702811  | 19.257 ug/ml |
| 12) N-OCTACOSANE              | 16.892 | 2689271  | 19.827 ug/ml |
| -----                         |        |          |              |

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_F\Data\FF112520\  
Data File : FF007919.D  
Signal(s) : FID2B.ch  
Acq On : 25 Nov 2020 18:07  
Operator : DD\AJ  
Sample : L4880-06MSD  
Misc :  
ALS Vial : 63 Sample Multiplier: 1

Instrument :  
FID\_F  
ClientSampleId :  
KENS-B4-112320-0-2MSD

Integration File: autoint1.e  
Quant Time: Nov 26 00:33:32 2020  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_F\Method\FF110220.M  
Quant Title :  
QLast Update : Tue Nov 03 01:29:25 2020  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1uL  
Signal Phase : Rxi-1ms  
Signal Info : 20mx0.18mmx0.18um

