

rteres

Instrument :
FID_F
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Rahul Chavli 11/26/2025
Supervised By :mohammad ahmed 11/27/2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF11252
Data File : FF016791.D
Signal(s) : FID2B.ch
Acq On : 25 Nov 2025 11:47
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF102925.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	4.563	4.530	4.667	BB	532398	5486093	96.27%	9.297%
2	6.742	6.700	6.845	BB	534234	5513803	96.76%	9.344%
3	8.575	8.537	8.684	BB	503057	5544564	97.30%	9.396%
4	10.185	10.149	10.289	BB	493971	5523849	96.93%	9.361%
5	11.629	11.590	11.724	BB	490629	5612405	98.49%	9.511%
6	12.942	12.900	13.029	BB	475238	5698591	100.00%	9.657%
7	14.143	14.102	14.234	BB	432208	5467905	95.95%	9.266%
8	15.045	15.004	15.114	BB	345812	4751011	83.37%	8.051%
9	15.250	15.209	15.322	BB	422326	5357749	94.02%	9.079%
10	16.274	16.230	16.349	BB	392209	5090745	89.33%	8.627%
11	17.226	17.184	17.285	M	371966	4962796	87.09%	8.410%
Sum of corrected areas:						59009511		

FF102925.M Wed Nov 26 02:53:42 2025