

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF112918\
 Data File : FF002843.D
 Signal(s) : FID2B.ch
 Acq On : 29 Nov 2018 13:07
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 29 14:09:06 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111418.M
 Quant Title :
 QLast Update : Wed Nov 14 13:46:00 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.650	3995022	51.142 ug/ml
Target Compounds			
1) N-OCTANE	2.277	4845820	38.581 ug/ml
2) N-DECANE	4.374	4914467	39.348 ug/ml
3) N-DODECANE	6.395	4826093	39.818 ug/ml
4) N-TETRADECANE	8.189	4734373	40.984 ug/ml
5) N-HEXADECANE	9.791	4698235	42.969 ug/ml
6) N-OCTADECANE	11.235	4688769	45.169 ug/ml
7) N-EICOSANE	12.548	4648809	48.013 ug/ml
8) N-DOCOSANE	13.748	4653926	50.831 ug/ml
10) N-TETRACOSANE	14.854	4658115	52.286 ug/ml
11) N-HEXACOSANE	15.879	4710435	51.904 ug/ml
12) N-OCTACOSANE	16.831	4852891	50.880 ug/ml
13) N-TRIACONTANE	17.721	5057741	51.018 ug/ml
14) N-DOTRIACONTANE	18.555	5274897	53.284 ug/ml
15) N-TETRATRIACONTANE	19.341	4912916	49.776 ug/ml
16) N-HEXATRIACONTANE	20.083	4105640	43.082 ug/ml
17) N-OCTATRIACONTANE	20.822	3536617	40.239 ug/ml
18) N-TETRACONTANE	21.722	2978549	34.004 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF112918\
Data File : FF002843.D
Signal(s) : FID2B.ch
Acq On : 29 Nov 2018 13:07
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 29 14:09:06 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111418.M
Quant Title :
QLast Update : Wed Nov 14 13:46:00 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

