

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF120220\
 Data File : FF007928.D
 Signal(s) : FID2B.ch
 Acq On : 02 Dec 2020 22:14
 Operator : DD\AJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 5 TRPH STD

Manual Integrations
 APPROVED

Ankita
 12/3/2020 10:25:22 AM

Integration File: autoint1.e
 Quant Time: Dec 03 00:26:01 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120220.M
 Quant Title :
 QLast Update : Thu Dec 03 00:20:32 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.708	624390	4.746 ug/ml
Target Compounds			
1) N-OCTANE	2.329	684049	4.617 ug/ml
2) N-DECANE	4.430	683803	4.555 ug/ml
3) N-DODECANE	6.453	685961	4.502 ug/ml
4) N-TETRADECANE	8.250	706251	4.565 ug/ml
5) N-HEXADECANE	9.852	714014	4.612 ug/ml
6) N-OCTADECANE	11.296	732475	4.653 ug/ml
7) N-EICOSANE	12.607	738115	4.718 ug/ml
8) N-DOCOSANE	13.807	721224	4.727 ug/ml
10) N-TETRACOSANE	14.912	720461	4.763 ug/ml
11) N-HEXACOSANE	15.935	696599	4.760 ug/ml
12) N-OCTACOSANE	16.887	671764	4.794 ug/ml
13) N-TRIACONTANE	17.778	635178	4.875 ug/ml
14) N-DOTRIACONTANE	18.612	566864	5.049 ug/ml
15) N-TETRATRIACONTANE	19.396	430698	4.908 ug/ml
16) N-HEXATRIACONTANE	20.138	334368	5.020 ug/mlm
17) N-OCTATRIACONTANE	20.888	261342	4.865 ug/ml
18) N-TETRACONTANE	21.823	315934	5.625 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF120220\
Data File : FF007928.D
Signal(s) : FID2B.ch
Acq On : 02 Dec 2020 22:14
Operator : DD\AJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 62 Sample Multiplier: 1

Instrument :
FID_F
ClientSampled :
5 TRPH STD

Manual Integrations
APPROVED

Ankita
12/3/2020 10:25:22 AM

Integration File: autoint1.e
Quant Time: Dec 03 00:26:01 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120220.M
Quant Title :
QLast Update : Thu Dec 03 00:20:32 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

