

rters

Instrument :
FID_F
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/05/2022
Supervised By :Ankita Jodhani 12/05/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF12022
Data File : FF011759.D
Signal(s) : FID2B.ch
Acq On : 02 Dec 2022 21:23
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 55 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF110422.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.089	2.045	2.165	BB	501855	5583997	82.60%	5.243%
2	4.615	4.578	4.663	BB	589995	6094884	90.16%	5.723%
3	6.778	6.743	6.816	BB	619172	6208226	91.83%	5.829%
4	8.604	8.568	8.643	BB	589985	6223379	92.06%	5.843%
5	10.210	10.174	10.258	BB	583537	6286001	92.98%	5.902%
6	11.653	11.613	11.691	BB	547080	6247098	92.41%	5.866%
7	12.962	12.916	13.008	BB	510379	6139254	90.81%	5.764%
8	14.161	14.084	14.208	BB	463017	5900261	87.28%	5.540%
9	15.059	15.011	15.131	BB	374532	5116102	75.68%	4.804%
10	15.264	15.213	15.311	BB	454085	5772195	85.38%	5.420%
11	16.285	16.233	16.340	BB	420796	5577549	82.50%	5.237%
12	17.234	17.157	17.276	BB	408772	5464623	80.83%	5.131%
13	18.121	18.062	18.159	BB	395416	5470328	80.92%	5.136%
14	18.954	18.890	18.995	BB	395889	5613576	83.04%	5.271%
15	19.738	19.689	19.776	M	417851	5880881	86.99%	5.522%
16	20.479	20.425	20.522	M	397489	5948219	87.99%	5.585%
17	21.221	21.169	21.272	M	350020	6215221	91.94%	5.836%
18	22.151	22.062	22.250	BB	275215	6760403	100.00%	6.348%
Sum of corrected areas:						106502200		

FF110422.M Sun Dec 04 11:54:36 2022