

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF120318\
 Data File : FF002886.D
 Signal(s) : FID2B.ch
 Acq On : 03 Dec 2018 20:40
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Dec 04 01:34:41 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111418.M
 Quant Title :
 QLast Update : Wed Nov 14 13:46:00 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.646	3619996	46.341 ug/ml
Target Compounds			
2) N-DECANE	4.370	4435079	35.509 ug/ml
3) N-DODECANE	6.391	4366957	36.030 ug/ml
4) N-TETRADECANE	8.185	4288812	37.127 ug/ml
5) N-HEXADECANE	9.786	4246531	38.838 ug/ml
6) N-OCTADECANE	11.230	4240186	40.848 ug/ml
7) N-EICOSANE	12.542	4220386	43.588 ug/ml
8) N-DOCOSANE	13.744	4220561	46.098 ug/ml
10) N-TETRACOSANE	14.849	4234456	47.531 ug/ml
11) N-HEXACOSANE	15.874	4286983	47.238 ug/ml
12) N-OCTACOSANE	16.827	4436351	46.512 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF120318\
Data File : FF002886.D
Signal(s) : FID2B.ch
Acq On : 03 Dec 2018 20:40
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Dec 04 01:34:41 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF111418.M
Quant Title :
QLast Update : Wed Nov 14 13:46:00 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

