

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF120518\
 Data File : FF002910.D
 Signal(s) : FID2B.ch
 Acq On : 05 Dec 2018 18:52
 Operator : AJ\SJ
 Sample : J6208-01MSD
 Misc :
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 TP-1MSD

Manual Integrations
 APPROVED

mohammad
 12/6/2018 11:58:01 AM

Integration File: autoint1.e
 Quant Time: Dec 06 03:06:37 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120418.M
 Quant Title :
 QLast Update : Tue Dec 04 16:40:56 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.650	2517658	19.393 ug/ml
Target Compounds			
1) N-OCTANE	2.270	1692446	10.941 ug/ml
2) N-DECANE	4.369	2590511	16.535 ug/ml
3) N-DODECANE	6.392	2843517	17.907 ug/ml
4) N-TETRADECANE	8.187	2991328	18.662 ug/ml
5) N-HEXADECANE	9.789	3162238	19.746 ug/ml
6) N-OCTADECANE	11.233	3077030	19.424 ug/mlm
7) N-EICOSANE	12.546	3057471	19.824 ug/ml
8) N-DOCOSANE	13.749	2987115	19.977 ug/ml
10) N-TETRACOSANE	14.854	2798255	19.363 ug/ml
11) N-HEXACOSANE	15.879	2706620	19.392 ug/ml
12) N-OCTACOSANE	16.828	2778750	20.739 ug/mlm
13) N-TRIACONTANE	17.719	2592325	20.231 ug/mlm
14) N-DOTRIACONTANE	18.556	2693073	22.551 ug/ml
15) N-TETRATRIACONTANE	19.342	2512361	22.168 ug/ml
16) N-HEXATRIACONTANE	20.083	2517898	23.336 ug/mlm
17) N-OCTATRIACONTANE	20.823	2354617	23.199 ug/ml
18) N-TETRACONTANE	21.727	2911255	27.699 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF120518\
Data File : FF002910.D
Signal(s) : FID2B.ch
Acq On : 05 Dec 2018 18:52
Operator : AJ\SJ
Sample : J6208-01MSD
Misc :
ALS Vial : 66 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
TP-1MSD

Manual Integrations
APPROVED

mohammad
12/6/2018 11:58:01 AM

Integration File: autoint1.e
Quant Time: Dec 06 03:06:37 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120418.M
Quant Title :
QLast Update : Tue Dec 04 16:40:56 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

