

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF120522\
 Data File : FF011768.D
 Signal(s) : FID2B.ch
 Acq On : 05 Dec 2022 19:32
 Operator : YP\AJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 5 TRPH STD

Integration File: autoint1.e
 Quant Time: Dec 06 04:18:09 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120522.M
 Quant Title :
 QLast Update : Tue Dec 06 04:16:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.056	612069	5.742 ug/ml
Target Compounds			
1) N-OCTANE	2.092	728706	5.807 ug/ml
2) N-DECANE	4.613	749169	5.800 ug/ml
3) N-DODECANE	6.775	748422	5.785 ug/ml
4) N-TETRADECANE	8.601	743811	5.775 ug/ml
5) N-HEXADECANE	10.207	749217	5.782 ug/ml
6) N-OCTADECANE	11.649	751330	5.779 ug/ml
7) N-EICOSANE	12.958	731051	5.784 ug/ml
8) N-DOCOSANE	14.157	714621	5.785 ug/ml
10) N-TETRACOSANE	15.260	719002	5.880 ug/ml
11) N-HEXACOSANE	16.281	682440	5.766 ug/ml
12) N-OCTACOSANE	17.231	665067	5.783 ug/ml
13) N-TRIACONTANE	18.119	656067	5.813 ug/ml
14) N-DOTRIACONTANE	18.951	659893	5.923 ug/ml
15) N-TETRATRIACONTANE	19.733	638858	5.828 ug/ml
16) N-HEXATRIACONTANE	20.473	627912	5.856 ug/ml
17) N-OCTATRIACONTANE	21.215	629125	5.629 ug/ml
18) N-TETRACONTANE	22.140	715589	6.103 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF120522\
Data File : FF011768.D
Signal(s) : FID2B.ch
Acq On : 05 Dec 2022 19:32
Operator : YP\AJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
5 TRPH STD

Integration File: autoint1.e
Quant Time: Dec 06 04:18:09 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120522.M
Quant Title :
QLast Update : Tue Dec 06 04:16:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

