

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF120719\
 Data File : FF006021.D
 Signal(s) : FID2B.ch
 Acq On : 06 Dec 2019 12:28
 Operator : DD\AJ
 Sample : 10 TRPH STD
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 10 TRPH STD

Manual Integrations
 APPROVED

mohammad
 12/12/2019 9:02:24 AM

Integration File: autoint1.e
 Quant Time: Dec 06 13:17:48 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120719.M
 Quant Title :
 QLast Update : Fri Dec 06 13:16:52 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.642	1463876	10.097 ug/ml
Target Compounds			
1) N-OCTANE	2.309	1580525	10.101 ug/ml
2) N-DECANE	4.401	1643435	10.090 ug/ml
3) N-DODECANE	6.417	1684781	10.105 ug/ml
4) N-TETRADECANE	8.209	1700258	10.122 ug/ml
5) N-HEXADECANE	9.806	1730355	10.142 ug/ml
6) N-OCTADECANE	11.244	1718953	10.113 ug/ml
7) N-EICOSANE	12.550	1735812	10.157 ug/ml
8) N-DOCOSANE	13.745	1710995	10.162 ug/ml
10) N-TETRACOSANE	14.846	1677735	10.153 ug/ml
11) N-HEXACOSANE	15.863	1615820	10.100 ug/ml
12) N-OCTACOSANE	16.811	1539278	10.093 ug/ml
13) N-TRIACONTANE	17.696	1461594	10.201 ug/ml
14) N-DOTRIACONTANE	18.527	1443472	11.103 ug/ml
15) N-TETRATRIACONTANE	19.311	1395651	11.700 ug/ml
16) N-HEXATRIACONTANE	20.048	1366717	12.629 ug/mlm
17) N-OCTATRIACONTANE	20.773	1299469	12.840 ug/ml
18) N-TETRACONTANE	21.663	1373032	13.478 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF120719\
Data File : FF006021.D
Signal(s) : FID2B.ch
Acq On : 06 Dec 2019 12:28
Operator : DD\AJ
Sample : 10 TRPH STD
Misc :
ALS Vial : 56 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
10 TRPH STD

Manual Integrations
APPROVED

mohammad
12/12/2019 9:02:24 AM

Integration File: autoint1.e
Quant Time: Dec 06 13:17:48 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120719.M
Quant Title :
QLast Update : Fri Dec 06 13:16:52 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

