

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF120719\
 Data File : FF006022.D
 Signal(s) : FID2B.ch
 Acq On : 06 Dec 2019 12:58
 Operator : DD\AJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 5 TRPH STD

Manual Integrations
 APPROVED

mohammad
 12/12/2019 9:02:27 AM

Integration File: autoint1.e
 Quant Time: Dec 06 13:17:57 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120719.M
 Quant Title :
 QLast Update : Fri Dec 06 13:16:52 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.642	729664	5.033 ug/ml
Target Compounds			
1) N-OCTANE	2.309	811227	5.184 ug/ml
2) N-DECANE	4.401	842176	5.170 ug/ml
3) N-DODECANE	6.417	859484	5.155 ug/ml
4) N-TETRADECANE	8.208	869862	5.178 ug/ml
5) N-HEXADECANE	9.806	883533	5.178 ug/ml
6) N-OCTADECANE	11.243	880425	5.180 ug/ml
7) N-EICOSANE	12.550	886924	5.190 ug/ml
8) N-DOCOSANE	13.745	875007	5.197 ug/ml
10) N-TETRACOSANE	14.845	856547	5.183 ug/ml
11) N-HEXACOSANE	15.864	823237	5.146 ug/ml
12) N-OCTACOSANE	16.812	780868	5.120 ug/ml
13) N-TRIACONTANE	17.697	732290	5.111 ug/ml
14) N-DOTRIACONTANE	18.528	704192	5.417 ug/ml
15) N-TETRATRIACONTANE	19.311	608756	5.103 ug/ml
16) N-HEXATRIACONTANE	20.049	566739	5.237 ug/mlm
17) N-OCTATRIACONTANE	20.776	507117	5.011 ug/ml
18) N-TETRACONTANE	21.663	607331	5.962 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF120719\
Data File : FF006022.D
Signal(s) : FID2B.ch
Acq On : 06 Dec 2019 12:58
Operator : DD\AJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
FID_F
ClientSampled :
5 TRPH STD

Manual Integrations
APPROVED

mohammad
12/12/2019 9:02:27 AM

Integration File: autoint1.e
Quant Time: Dec 06 13:17:57 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120719.M
Quant Title :
QLast Update : Fri Dec 06 13:16:52 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

