

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF120921\
 Data File : FF009335.D
 Signal(s) : FID2B.ch
 Acq On : 09 Dec 2021 15:13
 Operator : AJ\MA
 Sample : PB141269BS
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Dec 09 21:17:01 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF112321.M
 Quant Title :
 QLast Update : Wed Nov 24 00:27:43 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.682	2302367	19.058 ug/ml
Target Compounds			
2) N-DECANE	4.431	2944256	21.492 ug/ml
3) N-DODECANE	6.447	3015442	21.691 ug/ml
4) N-TETRADECANE	8.241	3104937	21.849 ug/ml
5) N-HEXADECANE	9.838	3008504	21.312 ug/ml
6) N-OCTADECANE	11.278	3090693	21.623 ug/ml
7) N-EICOSANE	12.586	3014383	21.174 ug/ml
8) N-DOCOSANE	13.783	2864936	20.668 ug/ml
10) N-TETRACOSANE	14.884	2772338	20.260 ug/ml
11) N-HEXACOSANE	15.905	2586045	19.553 ug/ml
12) N-OCTACOSANE	16.854	2398444	18.830 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF120921\
Data File : FF009335.D
Signal(s) : FID2B.ch
Acq On : 09 Dec 2021 15:13
Operator : AJ\MA
Sample : PB141269BS
Misc :
ALS Vial : 56 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Dec 09 21:17:01 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF112321.M
Quant Title :
QLast Update : Wed Nov 24 00:27:43 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

