

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF121020\
 Data File : FF007955.D
 Signal(s) : FID2B.ch
 Acq On : 10 Dec 2020 18:15
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Dec 11 01:02:58 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120220.M
 Quant Title :
 QLast Update : Thu Dec 03 09:08:10 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.708	5357841	42.379 ug/ml
Target Compounds			
2) N-DECANE	4.429	6825810	47.935 ug/ml
3) N-DODECANE	6.453	6829107	47.420 ug/ml
4) N-TETRADECANE	8.249	6832343	46.546 ug/ml
5) N-HEXADECANE	9.852	6726669	45.602 ug/ml
6) N-OCTADECANE	11.295	6720836	44.713 ug/ml
7) N-EICOSANE	12.606	6555349	43.706 ug/ml
8) N-DOCOSANE	13.807	6273082	42.844 ug/ml
10) N-TETRACOSANE	14.913	6105969	41.928 ug/ml
11) N-HEXACOSANE	15.936	5846867	41.491 ug/ml
12) N-OCTACOSANE	16.888	5666689	41.872 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF121020\
Data File : FF007955.D
Signal(s) : FID2B.ch
Acq On : 10 Dec 2020 18:15
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Dec 11 01:02:58 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120220.M
Quant Title :
QLast Update : Thu Dec 03 09:08:10 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

