

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF121021\
 Data File : FF009351.D
 Signal(s) : FID2B.ch
 Acq On : 10 Dec 2021 17:13
 Operator : AJ\MA
 Sample : PB141322BS
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 PB141322BS

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 12/11/2021
 Supervised By :Sohil Jodhani 12/22/2021

Integration File: autoint1.e
 Quant Time: Dec 10 23:11:43 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF112321.M
 Quant Title :
 QLast Update : Wed Nov 24 00:27:43 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.681	2458086	20.347 ug/ml
Target Compounds			
1) N-OCTANE	2.340	2388427	17.437 ug/ml
2) N-DECANE	4.431	2599213	18.973 ug/ml
3) N-DODECANE	6.447	2720763	19.572 ug/ml
4) N-TETRADECANE	8.240	2801232	19.712 ug/ml
5) N-HEXADECANE	9.838	2742138	19.425 ug/ml
6) N-OCTADECANE	11.279	2842090	19.884 ug/ml
7) N-EICOSANE	12.586	2784703	19.560 ug/ml
8) N-DOCOSANE	13.783	2698757	19.470 ug/ml
10) N-TETRACOSANE	14.885	2561033	18.716 ug/ml
11) N-HEXACOSANE	15.905	2380407	17.998 ug/ml
12) N-OCTACOSANE	16.852	2203426	17.299 ug/ml
13) N-TRIACONTANE	17.737	1987787	16.237 ug/ml
14) N-DOTRIACONTANE	18.567	1662467	14.476 ug/ml
15) N-TETRATRIACONTANE	19.349	1205256	11.225 ug/ml
16) N-HEXATRIACONTANE	20.088	900669	8.991 ug/mlm
17) N-OCTATRIACONTANE	20.821	636505	6.889 ug/mlm
18) N-TETRACONTANE	21.729	456831	4.886 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF121021\
Data File : FF009351.D
Signal(s) : FID2B.ch
Acq On : 10 Dec 2021 17:13
Operator : AJ\MA
Sample : PB141322BS
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
PB141322BS

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 12/11/2021
Supervised By :Sohil Jodhani 12/22/2021

Integration File: autoint1.e
Quant Time: Dec 10 23:11:43 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF112321.M
Quant Title :
QLast Update : Wed Nov 24 00:27:43 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

