

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF121118\
Data File : FF002976.D
Signal(s) : FID2B.ch
Acq On : 11 Dec 2018 10:59
Operator : AJ\SJ
Sample : RT MARKER
Misc :
ALS Vial : 54 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Dec 12 02:49:43 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120418.M
Quant Title :
QLast Update : Tue Dec 04 16:40:56 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
Target Compounds			
3) N-DODECANE	6.387	3669756	23.110 ug/ml
10) N-TETRACOSANE	14.820	106874	0.740 ug/ml
12) N-OCTACOSANE	16.822	3357198	25.056 ug/ml
13) N-TRIACONTANE	17.709	326695	2.550 ug/ml
14) N-DOTRIACONTANE	18.545	436784	3.658 ug/ml
15) N-TETRATRIACONTANE	19.332	320498	2.828 ug/ml
16) N-HEXATRIACONTANE	20.076	188707	1.749 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF121118\
Data File : FF002976.D
Signal(s) : FID2B.ch
Acq On : 11 Dec 2018 10:59
Operator : AJ\SJ
Sample : RT MARKER
Misc :
ALS Vial : 54 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Dec 12 02:49:43 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120418.M
Quant Title :
QLast Update : Tue Dec 04 16:40:56 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#3 N-DODECANE

R.T.: 6.387 min
Delta R.T.: -0.004 min
Response: 3669756
Conc: 23.11 ug/ml




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 0.000 min
Exp R.T. : 14.645 min
Response: 0
Conc: N.D.




#10 N-TETRACOSANE

R.T.: 14.820 min
Delta R.T.: -0.030 min
Response: 106874
Conc: 0.74 ug/ml




#12 N-OCTACOSANE

R.T.: 16.822 min
Delta R.T.: -0.003 min
Response: 3357198
Conc: 25.06 ug/ml




#13 N-TRIACONTANE

R.T.: 17.709 min
Delta R.T.: -0.006 min
Response: 326695
Conc: 2.55 ug/ml




#14 N-DOTRIACONTANE

R.T.: 18.545 min
Delta R.T.: -0.004 min
Response: 436784
Conc: 3.66 ug/ml



#15 N-TETRATRIACONTANE

R.T.: 19.332 min
Delta R.T.: -0.004 min
Response: 320498
Conc: 2.83 ug/ml



#16 N-HEXATRIACONTANE

R.T.: 20.076 min
Delta R.T.: -0.001 min
Response: 188707
Conc: 1.75 ug/ml