

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF121118\
 Data File : FF002980.D
 Signal(s) : FID2B.ch
 Acq On : 11 Dec 2018 13:02
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: Dec 12 02:50:51 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120418.M
 Quant Title :
 QLast Update : Tue Dec 04 16:40:56 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.644	5395842	41.562 ug/ml
Target Compounds			
1) N-OCTANE	2.265	7345414	47.483 ug/ml
2) N-DECANE	4.366	7391957	47.182 ug/ml
3) N-DODECANE	6.389	7307254	46.016 ug/ml
4) N-TETRADECANE	8.183	7136257	44.522 ug/ml
5) N-HEXADECANE	9.785	6961590	43.470 ug/ml
6) N-OCTADECANE	11.229	6772822	42.753 ug/ml
7) N-EICOSANE	12.540	6483197	42.036 ug/ml
8) N-DOCOSANE	13.742	6230798	41.669 ug/ml
10) N-TETRACOSANE	14.848	6129339	42.413 ug/ml
11) N-HEXACOSANE	15.873	5875996	42.098 ug/ml
12) N-OCTACOSANE	16.825	5833628	43.539 ug/ml
13) N-TRIACONTANE	17.714	5876043	45.859 ug/ml
14) N-DOTRIACONTANE	18.549	5848851	48.977 ug/ml
15) N-TETRATRIACONTANE	19.338	5917067	52.209 ug/ml
16) N-HEXATRIACONTANE	20.082	5833330	54.064 ug/ml
17) N-OCTATRIACONTANE	20.823	5459237	53.788 ug/ml
18) N-TETRACONTANE	21.728	5560183	52.901 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF121118\
Data File : FF002980.D
Signal(s) : FID2B.ch
Acq On : 11 Dec 2018 13:02
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: Dec 12 02:50:51 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120418.M
Quant Title :
QLast Update : Tue Dec 04 16:40:56 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

