

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF121421AL\
 Data File : FF009434.D
 Signal(s) : FID2B.ch
 Acq On : 15 Dec 2021 15:20
 Operator : AJ\MA
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: Dec 16 00:14:40 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\Aliphatic EPH -FF121421.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 14 16:37:50 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	10.735	2612449	19.434 ug/ml
Spiked Amount 50.000		Recovery =	38.87%
12) S 1-chlorooctadecane (S...	12.169	2066687	19.340 ug/ml
Spiked Amount 50.000		Recovery =	38.68%
Target Compounds			
1) T n-Nonane (C9)	2.636	2326532	20.062 ug/ml
2) T n-Decane (C10)	3.636	2361049	20.373 ug/ml
3) T A~Naphthalene (C11.7)	5.186	2625297	20.863 ug/ml
4) T n-Dodecane (C12)	5.613	2376095	20.204 ug/ml
5) T A~2-methylnaphthalene...	6.230	2663253	21.249 ug/ml
6) T n-Tetradecane (C14)	7.396	2408894	20.382 ug/ml
7) T n-Hexadecane (C16)	8.988	2427989	20.221 ug/ml
8) T n-Octadecane (C18)	10.423	2481473	19.639 ug/ml
10) T n-Eicosane (C20)	11.727	2445528	19.149 ug/ml
11) T n-Heneicosane (C21)	12.336	2450570	18.866 ug/ml
13) T n-Docosane (C22)	12.921	2430799	18.759 ug/ml
14) T n-Tetracosane (C24)	14.020	2416797	18.553 ug/ml
15) T n-Hexacosane (C26)	15.036	2321089	18.658 ug/ml
16) T n-Octacosane (C28)	15.982	2252592	19.481 ug/ml
17) T n-Tricontane (C30)	16.867	2170382	20.310 ug/ml
18) T n-Dotriacontane (C32)	17.696	2070636	20.110 ug/ml
19) T n-Tetratriacontane (C34)	18.475	2034479	19.845 ug/ml
20) T n-Hexatriacontane (C36)	19.213	1967215	19.663 ug/ml
21) T n-Octatriacontane (C38)	19.920	1956808	20.206 ug/ml
22) T n-Tetracontane (C40)	20.750	2020299	24.828 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF121421AL\
Data File : FF009434.D
Signal(s) : FID2B.ch
Acq On : 15 Dec 2021 15:20
Operator : AJ\MA
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 54 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
Quant Time: Dec 16 00:14:40 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\Aliphatic EPH -FF121421.M
Quant Title : GC Extractables
QLast Update : Tue Dec 14 16:37:50 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

