

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF121819\
 Data File : FF006046.D
 Signal(s) : FID2B.ch
 Acq On : 18 Dec 2019 16:22
 Operator : DD\AJ
 Sample : K6300-03MSD
 Misc :
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 SSI-MW-4MSD

Integration File: autoint1.e
 Quant Time: Dec 19 00:28:54 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120719.M
 Quant Title :
 QLast Update : Fri Dec 06 13:20:24 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.632	2459614	16.930 ug/ml
Target Compounds			
2) N-DECANE	4.395	2428113	14.788 ug/ml
3) N-DODECANE	6.411	2590152	15.409 ug/ml
4) N-TETRADECANE	8.201	2712943	15.999 ug/ml
5) N-HEXADECANE	9.797	2815795	16.342 ug/ml
6) N-OCTADECANE	11.235	2820511	16.451 ug/ml
7) N-EICOSANE	12.541	2797939	16.200 ug/ml
8) N-DOCOSANE	13.735	2744645	16.124 ug/ml
10) N-TETRACOSANE	14.835	2672211	16.011 ug/ml
11) N-HEXACOSANE	15.853	2557110	15.879 ug/ml
12) N-OCTACOSANE	16.799	2454054	15.998 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF121819\
Data File : FF006046.D
Signal(s) : FID2B.ch
Acq On : 18 Dec 2019 16:22
Operator : DD\AJ
Sample : K6300-03MSD
Misc :
ALS Vial : 60 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
SSI-MW-4MSD

Integration File: autoint1.e
Quant Time: Dec 19 00:28:54 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120719.M
Quant Title :
QLast Update : Fri Dec 06 13:20:24 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

