

Instrument :
FID_F
LabSampleId :
50 PPM TRPH STD

nteres

Area Percent Report

Manual IntegrationsAPPROVED

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF
Data File : FF013894.D
Signal(s) : FID2B.ch
Acq On : 18 Dec 2023 12:08
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Reviewed By :Yogesh Patel 12/19/2023
Supervised By :Ankita Jodhani 12/19/2023

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120523.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	1.996	1.951	2.081	BB	432960	5072847	74.54%	4.731%
2	4.517	4.473	4.572	BB	612096	6412345	94.22%	5.980%
3	6.683	6.639	6.742	BB	644204	6597278	96.93%	6.152%
4	8.510	8.446	8.572	BB	629768	6661639	97.88%	6.212%
5	10.117	10.071	10.185	BB	620426	6805881	100.00%	6.347%
6	11.560	11.513	11.622	BB	592264	6795937	99.85%	6.337%
7	12.871	12.818	12.944	BB	559487	6735371	98.96%	6.281%
8	14.071	14.036	14.138	VB	523167	6455912	94.86%	6.020%
9	14.972	14.898	14.999	BV	430328	6025160	88.53%	5.619%
10	15.176	15.141	15.243	VB	498995	6371630	93.62%	5.942%
11	16.199	16.143	16.271	BB	456294	6213551	91.30%	5.794%
12	17.151	17.070	17.227	BB	468950	6198747	91.08%	5.780%
13	18.039	17.961	18.097	BB	447326	6400427	94.04%	5.969%
14	18.873	18.806	18.943	BB	448329	6348182	93.27%	5.920%
15	19.657	19.589	19.695	PV	366527	5476797	80.47%	5.107%
16	20.397	20.345	20.440	BBA	290774	4404805	64.72%	4.108%
17	21.128	21.050	21.199	BV	239486	4084138	60.01%	3.809%
18	22.021	21.837	22.174	BB	173296	4175758	61.36%	3.894%

Sum of corrected areas: 107236404

FF120523.M Tue Dec 19 00:04:15 2023