

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF122718\
Data File : FF003098.D
Signal(s) : FID2B.ch
Acq On : 27 Dec 2018 19:10
Operator : AJ\SJ
Sample : RT MARKER
Misc :
ALS Vial : 54 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Dec 27 23:57:18 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120418.M
Quant Title :
QLast Update : Tue Dec 04 16:40:56 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
Target Compounds			
3) N-DODECANE	6.382	4962165	31.249 ug/ml
12) N-OCTACOSANE	16.810	4352493	32.485 ug/ml
13) N-TRIACONTANE	17.698	486826	3.799 ug/ml
14) N-DOTRIACONTANE	18.533	585365	4.902 ug/ml
15) N-TETRATRIACONTANE	19.321	430694	3.800 ug/ml
16) N-HEXATRIACONTANE	20.067	284508	2.637 ug/ml
17) N-OCTATRIACONTANE	20.804	192840	1.900 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF122718\
Data File : FF003098.D
Signal(s) : FID2B.ch
Acq On : 27 Dec 2018 19:10
Operator : AJ\SJ
Sample : RT MARKER
Misc :
ALS Vial : 54 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Dec 27 23:57:18 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120418.M
Quant Title :
QLast Update : Tue Dec 04 16:40:56 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#3 N-DODECANE

R.T.: 6.382 min
Delta R.T.: -0.009 min
Response: 4962165
Conc: 31.25 ug/ml




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 0.000 min
Exp R.T. : 14.645 min
Response: 0
Conc: N.D.




#12 N-OCTACOSANE

R.T.: 16.810 min
Delta R.T.: -0.015 min
Response: 4352493
Conc: 32.48 ug/ml




#13 N-TRIACONTANE

R.T.: 17.698 min
Delta R.T.: -0.017 min
Response: 486826
Conc: 3.80 ug/ml




#14 N-DOTRIACONTANE

R.T.: 18.533 min
Delta R.T.: -0.015 min
Response: 585365
Conc: 4.90 ug/ml




#15 N-TETRATRIACONTANE

R.T.: 19.321 min
Delta R.T.: -0.014 min
Response: 430694
Conc: 3.80 ug/ml



#16 N-HEXATRIACONTANE

R.T.: 20.067 min
Delta R.T.: -0.010 min
Response: 284508
Conc: 2.64 ug/ml



#17 N-OCTATRIACONTANE

R.T.: 20.804 min
Delta R.T.: -0.012 min
Response: 192840
Conc: 1.90 ug/ml