

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF122718\
 Data File : FF003095.D
 Signal(s) : FID2B.ch
 Acq On : 27 Dec 2018 17:40
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Dec 27 23:56:26 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120418.M
 Quant Title :
 QLast Update : Tue Dec 04 16:40:56 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.634	6169613	47.522 ug/ml
Target Compounds			
2) N-DECANE	4.361	8090075	51.638 ug/ml
3) N-DODECANE	6.384	7992486	50.332 ug/ml
4) N-TETRADECANE	8.177	7841771	48.923 ug/ml
5) N-HEXADECANE	9.778	7696882	48.062 ug/ml
6) N-OCTADECANE	11.221	7550724	47.664 ug/ml
7) N-EICOSANE	12.532	7288059	47.255 ug/ml
8) N-DOCOSANE	13.733	7062400	47.231 ug/ml
10) N-TETRACOSANE	14.838	6937053	48.002 ug/ml
11) N-HEXACOSANE	15.862	6766121	48.476 ug/ml
12) N-OCTACOSANE	16.815	6766623	50.502 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF122718\
Data File : FF003095.D
Signal(s) : FID2B.ch
Acq On : 27 Dec 2018 17:40
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Dec 27 23:56:26 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120418.M
Quant Title :
QLast Update : Tue Dec 04 16:40:56 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

