

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF122821\
 Data File : FF009572.D
 Signal(s) : FID2B.ch
 Acq On : 28 Dec 2021 10:49
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/29/2021
 Supervised By :Ankita Jodhani 12/29/2021

Integration File: autoint1.e
 Quant Time: Dec 29 00:37:32 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF121521.M
 Quant Title :
 QLast Update : Thu Dec 16 03:33:07 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.472	5501614	47.493 ug/ml
Target Compounds			
2) N-DECANE	4.256	7976407	55.944 ug/mlm
3) N-DODECANE	6.266	7863187	54.373 ug/ml
4) N-TETRADECANE	8.051	7661607	52.696 ug/ml
5) N-HEXADECANE	9.644	7319812	50.967 ug/ml
6) N-OCTADECANE	11.080	7142825	49.489 ug/ml
7) N-EICOSANE	12.384	6849125	48.328 ug/ml
8) N-DOCOSANE	13.576	6460757	47.413 ug/ml
10) N-TETRACOSANE	14.675	6252494	47.487 ug/ml
11) N-HEXACOSANE	15.692	6004162	48.040 ug/ml
12) N-OCTACOSANE	16.638	5916926	49.690 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF122821\
Data File : FF009572.D
Signal(s) : FID2B.ch
Acq On : 28 Dec 2021 10:49
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/29/2021
Supervised By :Ankita Jodhani 12/29/2021

Integration File: autoint1.e
Quant Time: Dec 29 00:37:32 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF121521.M
Quant Title :
QLast Update : Thu Dec 16 03:33:07 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

