

rters

Instrument :
FID_F
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/29/2021
Supervised By :Ankita Jodhani 12/29/2021

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF12282
Data File : FF009579.D
Signal(s) : FID2B.ch
Acq On : 28 Dec 2021 15:52
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF121521.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.186	2.141	2.243	BB	1127174	7948456	98.08%	6.830%
2	4.255	4.222	4.315	BV	900588	8103735	100.00%	6.963%
3	6.264	6.224	6.323	BB	816457	7962423	98.26%	6.842%
4	8.050	8.013	8.109	BB	752354	7762583	95.79%	6.670%
5	9.644	9.603	9.688	BV	681447	7436584	91.77%	6.390%
6	11.079	11.029	11.135	BB	645508	7299192	90.07%	6.272%
7	12.383	12.340	12.429	BB	597263	7036033	86.82%	6.046%
8	13.576	13.539	13.636	BB	533953	6674810	82.37%	5.735%
9	14.471	14.424	14.540	BB	406503	5688809	70.20%	4.888%
10	14.675	14.580	14.736	BB	508368	6486331	80.04%	5.573%
11	15.691	15.615	15.735	BV	488257	6190800	76.39%	5.320%
12	16.637	16.553	16.698	BB	455177	6067030	74.87%	5.213%
13	17.521	17.466	17.565	BB	430555	5973568	73.71%	5.133%
14	18.352	18.287	18.394	BB	410820	5804402	71.63%	4.987%
15	19.130	19.076	19.178	BB	392467	5548718	68.47%	4.768%
16	19.868	19.806	19.916	BB	348240	5113888	63.11%	4.394%
17	20.572	20.530	20.629	BB	305278	4706095	58.07%	4.044%
18	21.399	21.317	21.474	BB	208722	4575604	56.46%	3.932%
Sum of corrected areas:						116379061		

FF121521.M Wed Dec 29 10:14:44 2021