

rters

Instrument :
FID_F
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/30/2021
Supervised By :Ankita Jodhani 12/30/2021

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF12292
Data File : FF009588.D
Signal(s) : FID2B.ch
Acq On : 29 Dec 2021 10:35
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF121521.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.186	2.155	2.293	PH	1137082	8022854	97.14%	7.101%
2	4.254	4.223	4.321	PH	912597	8258719	100.00%	7.309%
3	6.263	6.229	6.324	PH	829436	8099357	98.07%	7.168%
4	8.049	8.014	8.089	BH	745319	7837955	94.91%	6.937%
5	9.642	9.606	9.683	HH	682603	7436116	90.04%	6.581%
6	11.076	11.024	11.249	BH	650829	7213577	87.34%	6.384%
7	12.380	12.341	12.530	BH	588232	6917330	83.76%	6.122%
8	13.573	13.536	13.672	PH	531667	6462609	78.25%	5.720%
9	14.469	14.405	14.594	BH	414669	5534165	67.01%	4.898%
10	14.672	14.594	14.772	PH	502967	6296657	76.24%	5.573%
11	15.689	15.576	15.791	BH	467068	6095118	73.80%	5.394%
12	16.634	16.555	16.705	HH	442287	6080409	73.62%	5.381%
13	17.519	17.439	17.617	HH	432608	6166725	74.67%	5.458%
14	18.348	18.270	18.409	BH	416290	5782423	70.02%	5.118%
15	19.127	19.059	19.169	BH	373842	5353335	64.82%	4.738%
16	19.862	19.810	19.933	M	322235	4634565	56.12%	4.102%
17	20.569	20.517	20.635	M	238646	3821632	46.27%	3.382%
18	21.393	21.336	21.485	M	133980	2974748	36.02%	2.633%
Sum of corrected areas:						112988292		

FF121521.M Wed Dec 29 12:10:57 2021