

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF123120\
 Data File : FF008031.D
 Signal(s) : FID2B.ch
 Acq On : 31 Dec 2020 10:01
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 01 00:20:47 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120220.M
 Quant Title :
 QLast Update : Thu Dec 03 09:08:10 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.698	5935083	46.945 ug/ml
Target Compounds			
2) N-DECANE	4.422	7629694	53.580 ug/ml
3) N-DODECANE	6.446	7563159	52.518 ug/ml
4) N-TETRADECANE	8.242	7504788	51.127 ug/ml
5) N-HEXADECANE	9.843	7365077	49.930 ug/ml
6) N-OCTADECANE	11.286	7357177	48.947 ug/ml
7) N-EICOSANE	12.598	7197485	47.988 ug/ml
8) N-DOCOSANE	13.798	6920281	47.264 ug/ml
10) N-TETRACOSANE	14.904	6800559	46.697 ug/ml
11) N-HEXACOSANE	15.926	6563094	46.573 ug/ml
12) N-OCTACOSANE	16.878	6398756	47.281 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF123120\
Data File : FF008031.D
Signal(s) : FID2B.ch
Acq On : 31 Dec 2020 10:01
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 01 00:20:47 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120220.M
Quant Title :
QLast Update : Thu Dec 03 09:08:10 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

