

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF032219\
 Data File : FF003953.D
 Signal(s) : FID2B.ch
 Acq On : 22 Mar 2019 12:19
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Sohil
 3/25/2019 4:25:35 PM

Integration File: autoint1.e
 Quant Time: Mar 22 22:58:09 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF030119.M
 Quant Title :
 QLast Update : Mon Mar 04 08:35:38 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.582	6870839	49.350 ug/ml
Target Compounds			
1) N-OCTANE	2.221	8437406	54.959 ug/ml
2) N-DECANE	4.311	9068887	58.958 ug/ml
3) N-DODECANE	6.333	9111268	58.382 ug/ml
4) N-TETRADECANE	8.125	9032064	56.333 ug/ml
5) N-HEXADECANE	9.726	8861705	54.410 ug/ml
6) N-OCTADECANE	11.168	8629838	52.380 ug/ml
7) N-EICOSANE	12.480	8282687	50.630 ug/ml
8) N-DOCOSANE	13.680	8000861	49.575 ug/ml
10) N-TETRACOSANE	14.785	7787795	49.221 ug/ml
11) N-HEXACOSANE	15.809	7697060	49.870 ug/ml
12) N-OCTACOSANE	16.762	7829780	51.814 ug/ml
13) N-TRIACONTANE	17.652	8079120	54.923 ug/ml
14) N-DOTRIACONTANE	18.489	8278678	58.214 ug/ml
15) N-TETRATRIACONTANE	19.273	8015178	59.861 ug/ml
16) N-HEXATRIACONTANE	20.016	7981813	61.172 ug/mlm
17) N-OCTATRIACONTANE	20.748	7242443	62.725 ug/mlm
18) N-TETRACONTANE	21.627	8850103	71.138 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF032219\
Data File : FF03953.D
Signal(s) : FID2B.ch
Acq On : 22 Mar 2019 12:19
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampled :
50 PPM TRPH STD

Manual Integrations
APPROVED

Sohil
3/25/2019 4:25:35 PM

Integration File: autoint1.e
Quant Time: Mar 22 22:58:09 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF030119.M
Quant Title :
QLast Update : Mon Mar 04 08:35:38 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rx1-1ms
Signal Info : 20mx0.18mmx0.18um

