

Method Path : P:\HPCHEM1\FID_F\Method\
 Method File : FF020118.M
 Title :
 Last Update : Fri Feb 02 10:05:00 2018
 Response Via : Initial Calibration

Calibration Files

50 =FF000005.D 100 =FF000004.D 20 =FF000006.D
 10 =FF000007.D 5 =FF000008.D

	Compound	50	100	20	10	5	Avg		%RSD
1)	N-OCTANE	1.504	1.459	1.570	1.611	1.425	1.514	E5	5.07
2)	N-DECANE	1.535	1.491	1.607	1.651	1.459	1.549	E5	5.15
3)	N-DODECANE	1.527	1.475	1.604	1.653	1.464	1.544	E5	5.33
4)	N-TETRADECANE	1.507	1.435	1.594	1.647	1.452	1.527	E5	5.98
5)	N-HEXADECANE	1.451	1.376	1.542	1.580	1.404	1.471	E5	5.98
6)	N-OCTADECANE	1.417	1.332	1.523	1.572	1.431	1.455	E5	6.47
7)	N-EICOSANE	1.343	1.275	1.435	1.462	1.314	1.366	E5	5.83
8)	N-DOCOSANE	1.290	1.223	1.373	1.386	1.244	1.303	E5	5.67
9) S	TETRACOSANE-d50 (S...	1.121	1.073	1.194	1.207	1.100	1.139	E5	5.17
10)	N-TETRACOSANE	1.250	1.197	1.329	1.339	1.220	1.267	E5	5.05
11)	N-HEXACOSANE	1.201	1.159	1.267	1.272	1.163	1.213	E5	4.53
12)	N-OCTACOSANE	1.146	1.121	1.205	1.219	1.116	1.162	E5	4.13
13)	N-TRIACONTANE	1.081	1.076	1.120	1.131	1.067	1.095	E5	2.61
14)	N-DOTRIACONTANE	0.956	0.982	0.991	1.016	0.986	0.986	E5	2.18
15)	N-TETRATRIACONTANE	7.880	8.300	8.639	8.362	8.116	8.259	E4	3.43
16)	N-HEXATRIACONTANE	6.153	6.581	6.875	6.911	6.251	6.554	E4	5.31
17)	N-OCTATRIACONTANE	5.500	5.894	6.458	6.695	5.664	6.042	E4	8.51
18)	N-TETRACONTANE	5.021	5.435	6.650	6.417	5.017	5.708	E4	13.61

(#) = Out of Range