

Data Path : P:\HPCHEM1\FID_F\Data\FF020118\
 Data File : FF000005.D
 Signal(s) : FID2B.ch
 Acq On : 01 Feb 2018 20:09
 Operator : UA\AJ
 Sample : 50 TRPH STD
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 02 10:02:25 2018
 Quant Method : P:\HPCHEM1\FID_F\Method\FF020118.M
 Quant Title :
 QLast Update : Fri Feb 02 10:02:11 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.636	5606037	50.000 ug/ml
Target Compounds			
1) N-OCTANE	2.287	7522082	50.000 ug/ml
2) N-DECANE	4.384	7676483	50.000 ug/ml
3) N-DODECANE	6.401	7634426	50.000 ug/ml
4) N-TETRADECANE	8.193	7534740	50.000 ug/ml
5) N-HEXADECANE	9.791	7257102	50.000 ug/ml
6) N-OCTADECANE	11.231	7084625	50.000 ug/ml
7) N-EICOSANE	12.539	6715646	50.000 ug/ml
8) N-DOCOSANE	13.737	6451298	50.000 ug/ml
10) N-TETRACOSANE	14.840	6251574	50.000 ug/ml
11) N-HEXACOSANE	15.860	6007204	50.000 ug/ml
12) N-OCTACOSANE	16.809	5731808	50.000 ug/ml
13) N-TRIACONTANE	17.697	5406532	50.000 ug/ml
14) N-DOTRIACONTANE	18.531	4780473	50.000 ug/ml
15) N-TETRATRIACONTANE	19.316	3940149	50.000 ug/ml
16) N-HEXATRIACONTANE	20.057	3076315	50.000 ug/ml
17) N-OCTATRIACONTANE	20.786	2750243	50.000 ug/ml
18) N-TETRACONTANE	21.670	2510438	50.000 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : P:\HPCHEM1\FID_F\Data\FF020118\
Data File : FF000005.D
Signal(s) : FID2B.ch
Acq On : 01 Feb 2018 20:09
Operator : UA\AJ
Sample : 50 TRPH STD
Misc :
ALS Vial : 54 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 02 10:02:25 2018
Quant Method : P:\HPCHEM1\FID_F\Method\FF020118.M
Quant Title :
QLast Update : Fri Feb 02 10:02:11 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#1 N-OCTANE

R.T.: 2.287 min
Delta R.T.: 0.000 min
Response: 7522082
Conc: 50.00 ug/ml




#2 N-DECANE

R.T.: 4.384 min
Delta R.T.: 0.000 min
Response: 7676483
Conc: 50.00 ug/ml




#3 N-DODECANE

R.T.: 6.401 min
Delta R.T.: 0.000 min
Response: 7634426
Conc: 50.00 ug/ml




#4 N-TETRADECANE

R.T.: 8.193 min
Delta R.T.: 0.000 min
Response: 7534740
Conc: 50.00 ug/ml




#5 N-HEXADECANE

R.T.: 9.791 min
Delta R.T.: 0.000 min
Response: 7257102
Conc: 50.00 ug/ml




#6 N-OCTADECANE

R.T.: 11.231 min
Delta R.T.: 0.000 min
Response: 7084625
Conc: 50.00 ug/ml




#7 N-EICOSANE

R.T.: 12.539 min
Delta R.T.: 0.000 min
Response: 6715646
Conc: 50.00 ug/ml




#8 N-DOCOSANE

R.T.: 13.737 min
Delta R.T.: 0.000 min
Response: 6451298
Conc: 50.00 ug/ml




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.636 min
Delta R.T.: 0.000 min
Response: 5606037
Conc: 50.00 ug/ml




#10 N-TETRACOSANE

R.T.: 14.840 min
Delta R.T.: 0.000 min
Response: 6251574
Conc: 50.00 ug/ml




#11 N-HEXACOSANE

R.T.: 15.860 min
Delta R.T.: 0.000 min
Response: 6007204
Conc: 50.00 ug/ml




#12 N-OCTACOSANE

R.T.: 16.809 min
Delta R.T.: 0.000 min
Response: 5731808
Conc: 50.00 ug/ml




#13 N-TRIACONTANE

R.T.: 17.697 min
Delta R.T.: 0.000 min
Response: 5406532
Conc: 50.00 ug/ml




#14 N-DOTRIACONTANE

R.T.: 18.531 min
Delta R.T.: 0.000 min
Response: 4780473
Conc: 50.00 ug/ml




#15 N-TETRATRIACONTANE

R.T.: 19.316 min
Delta R.T.: 0.000 min
Response: 3940149
Conc: 50.00 ug/ml




#16 N-HEXATRIACONTANE

R.T.: 20.057 min
Delta R.T.: 0.000 min
Response: 3076315
Conc: 50.00 ug/ml



#17 N-OCTATRIACONTANE

R.T.: 20.786 min
Delta R.T.: 0.000 min
Response: 2750243
Conc: 50.00 ug/ml



#18 N-TETRACONTANE

R.T.: 21.670 min
Delta R.T.: 0.000 min
Response: 2510438
Conc: 50.00 ug/ml