Quantitation Report (QT Reviewed)

Data Path : P:\HPCHEM1\FID_F\Data\FF021318\
Data File : FF000099.D
Signal(s) : FID2B.ch
Acq On : 13 Feb 2018 09:24
Operator : UA\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 13 11:47:14 2018
Quant Method : P:\HPCHEM1\FID_F\Method\FF020118.M
Quant Title :
QLast Update : Fri Feb 02 10:05:00 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

	Compound	R.T.	Response	Conc Units
	System Monitoring Compounds			
9) S	TETRACOSANE-d50 (SURR...	14.631	6064362	53.245 ug/ml
	Target Compounds			
1)	N-OCTANE	2.289	8252497	54.505 ug/ml
2)	N-DECANE	4.384	8696142	56.154 ug/ml
3)	N-DODECANE	6.400	8575309	55.523 ug/ml
4)	N-TETRADECANE	8.192	8389359	54.941 ug/ml
5)	N-HEXADECANE	9.790	8059057	54.796 ug/ml
6)	N-OCTADECANE	11.229	7774355	53.438 ug/ml
7)	N-EICOSANE	12.536	7354272	53.849 ug/ml
8)	N-DOCOSANE	13.733	6995423	53.675 ug/ml
10)	N-TETRACOSANE	14.835	6672227	52.653 ug/ml
11)	N-HEXACOSANE	15.854	6322501	52.139 ug/ml
12)	N-OCTACOSANE	16.803	6044721	52.039 ug/ml
13)	N-TRIACONTANE	17.690	5971506	54.529 ug/ml
14)	N-DOTRIACONTANE	18.522	5676032	57.567 ug/ml
15)	N-TETRATRIACONTANE	19.306	4810390	58.242 ug/ml
16)	N-HEXATRIACONTANE	20.046	3760375	57.374 ug/ml
17)	N-OCTATRIACONTANE	20.771	3166732	52.410 ug/ml
18)	N-TETRACONTANE	21.650	2139845	37.488 ug/ml

(f)=RT Delta > 1/2 Window (m)=manual int.

Data Path : P:\HPCHEM1\FID_F\Data\FF021318\
Data File : FF000099.D
Signal(s) : FID2B.ch
Acq On : 13 Feb 2018 09:24
Operator : UA\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 13 11:47:14 2018
Quant Method : P:\HPCHEM1\FID_F\Method\FF020118.M
Quant Title :
QLast Update : Fri Feb 02 10:05:00 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um