

Method Path : P:\HPCHEM1\FID_F\Method\
 Method File : FF042318.M
 Title :
 Last Update : Mon Apr 23 18:07:21 2018
 Response Via : Initial Calibration

Calibration Files

50 =FF000457.D 100 =FF000456.D 20 =FF000458.D
 10 =FF000459.D 5 =FF000460.D

	Compound	50	100	20	10	5	Avg	%RSD
1)	N-OCTANE	1.566	1.442	1.605	1.645	1.594	1.570 E5	4.91
2)	N-DECANE	1.594	1.462	1.630	1.669	1.607	1.593 E5	4.91
3)	N-DODECANE	1.614	1.479	1.649	1.686	1.613	1.608 E5	4.84
4)	N-TETRADECANE	1.631	1.493	1.668	1.706	1.626	1.625 E5	4.95
5)	N-HEXADECANE	1.645	1.505	1.688	1.727	1.641	1.641 E5	5.11
6)	N-OCTADECANE	1.652	1.511	1.699	1.740	1.654	1.651 E5	5.24
7)	N-EICOSANE	1.638	1.496	1.691	1.738	1.664	1.646 E5	5.54
8)	N-DOCOSANE	1.622	1.481	1.678	1.726	1.655	1.632 E5	5.68
9) S	TETRACOSANE-d50 (S...	1.423	1.299	1.477	1.518	1.451	1.434 E5	5.79
10)	N-TETRACOSANE	1.597	1.460	1.658	1.704	1.629	1.610 E5	5.76
11)	N-HEXACOSANE	1.574	1.440	1.639	1.681	1.612	1.589 E5	5.79
12)	N-OCTACOSANE	1.549	1.418	1.622	1.670	1.612	1.574 E5	6.18
13)	N-TRIACONTANE	1.524	1.396	1.605	1.664	1.612	1.560 E5	6.70
14)	N-DOTRIACONTANE	1.471	1.348	1.557	1.610	1.563	1.510 E5	6.85
15)	N-TETRATRIACONTANE	1.447	1.326	1.527	1.562	1.504	1.473 E5	6.24
16)	N-HEXATRIACONTANE	1.370	1.261	1.437	1.443	1.388	1.380 E5	5.33
17)	N-OCTATRIACONTANE	1.220	1.132	1.269	1.263	1.215	1.220 E5	4.51
18)	N-TETRACONTANE	1.191	1.118	1.235	1.228	1.191	1.193 E5	3.89

(#) = Out of Range