

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF072018\
 Data File : FF001447.D
 Signal(s) : FID2B.ch
 Acq On : 20 Jul 2018 18:49
 Operator : AJ\SJ
 Sample : J3762-07
 Misc : FID_F LOD WATER 1 PPB
 ALS Vial : 56 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 20 23:37:33 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_F\Method\FF070218.M
 Quant Title :
 QLast Update : Mon Jul 02 13:42:46 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.448	2604412	24.177 ug/ml
Target Compounds			
2) N-DECANE	4.211	81517	0.657 ug/mlm
3) N-DODECANE	6.225	77254	0.617 ug/mlm
4) N-TETRADECANE	8.034	89283	0.710 ug/mlm
5) N-HEXADECANE	9.620	88145	0.705 ug/mlm
6) N-OCTADECANE	11.054	92933	0.742 ug/mlm
7) N-EICOSANE	12.358	95729	0.770 ug/ml
8) N-DOCOSANE	13.551	101131	0.820 ug/ml
10) N-TETRACOSANE	14.651	101055	0.829 ug/ml
11) N-HEXACOSANE	15.675	100501	0.829 ug/ml
12) N-OCTACOSANE	16.620	115129	0.956 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF072018\
Data File : FF001447.D
Signal(s) : FID2B.ch
Acq On : 20 Jul 2018 18:49
Operator : AJ\SJ
Sample : J3762-07
Misc : FID_F LOD WATER 1 PPB
ALS Vial : 56 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 20 23:37:33 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_F\Method\FF070218.M
Quant Title :
QLast Update : Mon Jul 02 13:42:46 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

