

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF072018\
 Data File : FF001444.D
 Signal(s) : FID2B.ch
 Acq On : 20 Jul 2018 13:34
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 20 23:54:47 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_F\Method\FF070218.M
 Quant Title :
 QLast Update : Mon Jul 02 13:42:46 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.455	5882664	54.609 ug/ml
Target Compounds			
2) N-DECANE	4.219	7001215	56.390 ug/ml
3) N-DODECANE	6.232	7047075	56.240 ug/ml
4) N-TETRADECANE	8.020	7024749	55.855 ug/ml
5) N-HEXADECANE	9.617	6987256	55.923 ug/ml
6) N-OCTADECANE	11.056	6967131	55.594 ug/ml
7) N-EICOSANE	12.362	6859948	55.148 ug/ml
8) N-DOCOSANE	13.558	6783601	55.013 ug/ml
10) N-TETRACOSANE	14.659	6642134	54.485 ug/ml
11) N-HEXACOSANE	15.678	6545868	54.003 ug/ml
12) N-OCTACOSANE	16.626	6438323	53.442 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF072018\
Data File : FF001444.D
Signal(s) : FID2B.ch
Acq On : 20 Jul 2018 13:34
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 20 23:54:47 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_F\Method\FF070218.M
Quant Title :
QLast Update : Mon Jul 02 13:42:46 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#2 N-DECANE

R.T.: 4.219 min
Delta R.T.: -0.013 min
Response: 7001215
Conc: 56.39 ug/ml




#3 N-DODECANE

R.T.: 6.232 min
Delta R.T.: -0.012 min
Response: 7047075
Conc: 56.24 ug/ml




#4 N-TETRADECANE

R.T.: 8.020 min
Delta R.T.: -0.012 min
Response: 7024749
Conc: 55.85 ug/ml




#5 N-HEXADECANE

R.T.: 9.617 min
Delta R.T.: -0.011 min
Response: 6987256
Conc: 55.92 ug/ml




#6 N-OCTADECANE

R.T.: 11.056 min
Delta R.T.: -0.011 min
Response: 6967131
Conc: 55.59 ug/ml




#7 N-EICOSANE

R.T.: 12.362 min
Delta R.T.: -0.011 min
Response: 6859948
Conc: 55.15 ug/ml




#8 N-DOCOSANE

R.T.: 13.558 min
Delta R.T.: -0.010 min
Response: 6783601
Conc: 55.01 ug/ml




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.455 min
Delta R.T.: -0.012 min
Response: 5882664
Conc: 54.61 ug/ml




#10 N-TETRACOSANE

R.T.: 14.659 min
Delta R.T.: -0.012 min
Response: 6642134
Conc: 54.48 ug/ml



#11 N-HEXACOSANE

R.T.: 15.678 min
Delta R.T.: -0.011 min
Response: 6545868
Conc: 54.00 ug/ml



#12 N-OCTACOSANE

R.T.: 16.626 min
Delta R.T.: -0.011 min
Response: 6438323
Conc: 53.44 ug/ml