CHEMTECH

284 Sheffield Street Mountainside, NJ 07092

SDG NARRATIVE

USEPA SDG # MBFXD7 CASE # 48952 CONTRACT # EPW14030 SOW# ISM02.4 LAB NAME: CHEMTECH CONSULTING GROUP LAB CODE: CHM CHEMTECH PROJECT #L2885

A. Number of Samples and Date of Receipt

01 Water samples were delivered to the laboratory intact on 06/03/2020.

B. Parameters

Test requested for Metals CLP MS = Aluminum, Antimony, Arsenic, Barium, Beryllium, Cadmium, Calcium, Chromium, Cobalt, Copper, Iron, Lead, Magnesium, Manganese, Nickel, Potassium, Selenium, Silver, Sodium, Thallium, Vanadium, Zinc.

C. Cooler Temp

Indicator Bottle: <u>Presence/</u>Absence Cooler: 4.9°C

D. Detail Documentation (related to Sample Handling Shipping, Analytical Problem, Temp of Cooler etc):

Issue: Sample tags were not received with samples at the laboratory. Sample tag numbers may or may not be listed on the COC.

E. Corrective Action taken for above:

Resolution: In accordance with previous direction from [Region 07], the laboratory will note the issue in the SDG Narrative, and proceed with the analysis of the sample. The Resolution will be applied to all samples received for this Case.

F. Analytical Techniques:

All analyses were based on CLP Methodology by method ISM02.4.

<u>CHEMTECH</u> 284 Sheffield Street Mountainside, NJ 07092

G. Calculation:

Calculation for ICP-MS Water :

Concentration or Result ($\mu g/L$) = C x Vf Vi DF

Where,

C = Instrument value in ppb (The average of all replicate integrations)
Vf = Final digestion volume (mL)
Vi = Initial aliquot amount (mL) (Sample amount taken in prep)
DF = Dilution Factor

Example Calculation Sample For MBFXD7 for Zinc:

If C = 3.44 ppb Vf = 50 ml Vi = 50 ml DF = 1 Concentration or Result (μ g/L) = 3.44 x $\frac{50}{50}$ x 1 = 3.44 μ g/L = 3.4 μ g/L (Reported Result with Signification)

H. QA/QC

Calibrations met requirements. Interference check met requirements. Blank analyses did not indicate any presence of contamination. Laboratory Control sample was within control limits. Spike sample did meet requirements. Duplicate sample did meet requirements. Serial Dilution did meet requirements except for Nickel.

Chemical or physical interference effect was suspected and the data for all affected analytes in the sample received and associated with this serial dilution were flagged.

Collision cell is being used to remove potential interferences. The analytes Na, Mg, Al, K, V, Cr, Mn, Fe, Co, Ni, Cu, Zn, As are being analyzed with collision cell and analytes Be, B, Ca, Ti, Se, Sr, Zr, Mo, Ag, Cd, Sn, Sb, Ba, Tl, Pb, U are being analyzed with Non-Collision Cell. Helium gas is used for the Collision Cell analysis.

<u>CHEMTECH</u> 284 Sheffield Street Mountainside, NJ 07092

I certify that the data package is in compliance with the terms and conditions of the contract both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Director or his designee, as verified by the following signature.

Signature_____

Name: Nimisha Pandya

Date _____

Title: Document Control Officer