


LB108433

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.023	0.023	69	0.000	69				
0.2	0.200	0.195	-0.005	966	0.0 %	966				
2.5	2.500	2.485	-0.015	12886	0.0 %	12886				
5.0	5.000	4.961	-0.039	25775	0.0 %	25775				
7.5	7.500	7.554	0.054	39276	0.0 %	39276				
10.0	10.000	9.983	-0.017	51920	0.0 %	51920				

LB108433 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 Mar 2020 11:27:22

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	69	-	0.00007471B	PPB	25 Mar 2020 11:30:49	S	Std
0.2 - 1	50.2	966	-	0.20007471B	PPB	25 Mar 2020 11:33:05	S	Std
2.5 - 1	52.5	12886	-	2.50007471B	PPB	25 Mar 2020 11:35:23	S	Std
5.0 - 1	55	25775	-	5.00007471B	PPB	25 Mar 2020 11:37:39	S	Std
7.5 - 1	57.5	39276	-	7.50007471B	PPB	25 Mar 2020 11:39:56	S	Std
10.0 - 1	510	51920	-	10.00007471B	PPB	25 Mar 2020 11:44:50	S	Std
ICV75 - 1	ICV75	20708	3.9873	-7471B	PPB	25 Mar 2020 11:47:33	U	SMPL
ICB75 - 1	ICB75	-215	-0.0317	-7471B	PPB	25 Mar 2020 11:49:48	U	SMPL
CCV92 - 1	CCV92	26425	5.0854	-7471B	PPB	25 Mar 2020 11:52:05	U	SMPL
CCB92 - 1	CCB92	-228	-0.0342	-7471B	PPB	25 Mar 2020 11:54:20	U	SMPL
CRA - 1	CRA	951	0.1923	-7471B	PPB	25 Mar 2020 11:56:36	U	SMPL
HighStd - 1	HighStd	52987	10.1876	-7471B	PPB	25 Mar 2020 11:58:51	U	SMPL
ChkStd - 1	ChkStd	37291	7.1726	-7471B	PPB	25 Mar 2020 12:01:06	U	SMPL
PB127798BL - 1	PBS	-333	-0.0543	-7471B	PPB	25 Mar 2020 12:03:24	U	SMPL
PB127798BS - 1	LCSS	19482	3.7518	-7471B	PPB	25 Mar 2020 12:05:42	U	SMPL
L1754-01 - 1	NB-08-032420	3481	0.6783	-7471B	PPB	25 Mar 2020 12:07:58	U	SMPL
L1757-07 - 1	EO-01-032420	3082	0.6016	-7471B	PPB	25 Mar 2020 12:10:16	U	SMPL
L2017-01 - 1	FE-SB-32-(8.5-9)	2566	0.5025	-7471B	PPB	25 Mar 2020 12:12:33	U	SMPL
L2017-02 - 1	FE-SB-35-(7-7.5)	7763	1.5008	-7471B	PPB	25 Mar 2020 12:14:49	U	SMPL
L2017-03 - 1	FE-SB-29-(5.5-6)	536	0.1126	-7471B	PPB	25 Mar 2020 12:17:05	U	SMPL
CCV93 - 1	CCV93	27194	5.2332	-7471B	PPB	25 Mar 2020 12:19:20	U	SMPL
CCB93 - 1	CCB93	-108	-0.0111	-7471B	PPB	25 Mar 2020 12:21:35	U	SMPL
L2028-01 - 1	SB-30-(8-8.5)	1114	0.2236	-7471B	PPB	25 Mar 2020 12:26:42	U	SMPL
L2028-01DUP - 1	SB-30-(8-8.5)DUP	1031	0.2077	-7471B	PPB	25 Mar 2020 12:28:57	U	SMPL
L2028-01MS - 1	SB-30-(8-8.5)MS	21563	4.1515	-7471B	PPB	25 Mar 2020 12:31:12	U	SMPL
L2028-01MSD - 1	SB-30-(8-8.5)MSD	21294	4.0999	-7471B	PPB	25 Mar 2020 12:33:28	U	SMPL
L2028-02 - 1	SB-31-(5.5-6)	6248	1.2098	-7471B	PPB	25 Mar 2020 12:35:44	U	SMPL
L2028-01LX5 - 1		165	0.0413	-7471B	PPB	25 Mar 2020 12:38:01	U	SMPL
L2028-01A - 1		21389	4.1181	-7471B	PPB	25 Mar 2020 12:40:17	U	SMPL
CCV94 - 1	CCV94	27263	5.2464	-7471B	PPB	25 Mar 2020 12:42:33	U	SMPL
CCB94 - 1	CCB94	-95	-0.0086	-7471B	PPB	25 Mar 2020 12:44:51	U	SMPL