



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : sweta

Run Number: LB100007

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.3	4.01	12/21/2018	15:25
2	CAL2	1	Water			21.3	7.01	12/21/2018	15:26
3	CAL3	1	Water			21.3	10.02	12/21/2018	15:29
4	ICV	1	Water			21.3	6.99	12/21/2018	15:30
5	CCV1	1	Water			21.4	2.01	12/21/2018	15:31
6	J6539-01	1	Solid	20.03	20	22.6	7.51	12/21/2018	15:34
7	J6539-02	1	Solid	20.04	20	22.6	7.10	12/21/2018	15:35
8	J6539-03	1	Solid	20.02	20	22.6	7.19	12/21/2018	15:37
9	J6539-03DUP	1	Solid	20.03	20	22.5	7.20	12/21/2018	15:38
10	CCV2	1	Water			21.2	12.02	12/21/2018	15:40

CB 100007

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH122118

WorkList ID : 120726

Date : 12/21/2018 3:08:35 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/19/2018	Solid	J6539-01	pH	Cool 4 deg C	USEP01	A11	MYAJP1	12/19/2018	9045D
12/19/2018	Solid	J6539-02	pH	Cool 4 deg C	USEP01	A11	MYAJP2	12/19/2018	9045D
12/19/2018	Solid	J6539-03	pH	Cool 4 deg C	USEP01	A11	MYAJP3	12/19/2018	9045D

Date/Time 12.21.18 3:10 PM
Received by: JP
Relinquished by: SC

Date/Time 12.21.18 4:30 PM
Received by:
Relinquished by: JP